Anesthetic gas delivery systems

Compatibility with Hamilton Medical ventilators

The use of anesthetic gas delivery systems in combination with Hamilton Medical ventilators has been validated successfully as per the following table. Please follow the instructions for use in the operator's manual provided by the manufacturer of the respective product as shown below.

Manufacturer	System	Anesthetic gas removal	HAMILTON-S1	HAMILTON-G5	HAMILTON-C1	HAMILTON-T1	HAMILTON-MR1	HAMILTON-C2	HAMILTON-C3	HAMILTON-C6
Sedana	SEDACONDA 1)	FlurAbsorb 1-day FlurAbsorb 5-day	~	~	~	\checkmark	~	~	~	\checkmark
TIM	Mirus [§]	ORS	~	~	✓	✓	~	~	~	~

Compatible: 🗸 Not compatible: X

1) Please note that use of the SEDACONDA system may cause a slight increase in airway resistance.



EL020210121N.0

The products shown here are not available for purchase by the general public. The information provided here is intended for healthcare professionals only. Always read the labels and follow the product's instructions for use. Specifications are subject to change without notice. Some features are options. Not all features are available in all markets. All images are used for illustrative purposes only and may not accurately represent the product or its use. For all proprietary trademarks (®) and third-party trademarks used by Hamilton Medical AG see www.hamilton-medical.com/trademarks. ©2024 Hamilton Medical AG. All rights reserved.